

Certificate of Analysis

Product ID : OPU Tube
Product Code : AIT-14
Lot # : 202503
Expiry date : 2028-3
Storage Conditions : 3-30°C, Dry, No direct sunlight

Analysis	Ref#	Results	Date of Completion
Mouse Embryo Assay 80% of one cell stage control embryos to develop to expanded blastocyst within 96 hours	MEA.024.1578.2025	Pass	05-17-2025
Proven non-toxic by LAL test. Pass criteria <0.03 EU/device	LAL.024.1579.2025	Pass	05-09-2025
Sperm Motility Assay: >0.75 sperm motility by 8 hours post incubation and >0.8 sperm survival by 24 hours post incubation	SSMA.024.1580.2025	Pass	05-21-2025

Tests conducted by Embryotools

Quality Control Department
ASTEC CO., LTD.

