

Certificate of Analysis

Product ID : Sperm Collection Cup
Product Code : AIT-SC
Lot # : 202503
Expiry date : 2028-3
Storage Conditions : 3-30°C, Dry, No direct sunlight

Analysis	Ref#	Results	Date of Completion
Mouse Embryo Assay 80% of one cell stage control embryos to develop to expanded blastocyst within 96 hours	MEA.024.1584.2025	Pass	05-17-2025
Proven non-toxic by LAL test. Pass criteria <0.03 EU/device	LAL.024.1585.2025	Pass	05-09-2025
Sperm Motility Assay: >0.75 sperm motility by 8 hours post incubation and >0.8 sperm survival by 24 hours post incubation	SSMA.024.1586.2025	Pass	05-29-2025

Tests conducted by Embryotools

Quality Control Department
ASTEC CO., LTD.

