

Certificate of Analysis

Product ID : suregrip Center Well Dish
Product Code : AID-CW
Lot # : 241201
Expiry date : 2027-11-16
Storage Conditions : 3-30°C, Dry, No direct sunlight

Analysis	Ref#	Results	Date of Completion
Mouse Embryo Assay80% of one cell stage control embryos to develop to expanded blastocyst within 96 hours	MEA.024.124.2025	Pass	02-03-2025
Proven non-toxic by LAL test. Pass criteria <0.03 EU/device	LAL.024.125.2025	Pass	01-24-2025
Sperm Motility Assay: >0.75 sperm motility by 8 hours post incubation and >0.8 sperm survival by 24 hours post incubation	SSMA.024.126.2025	Pass	02-05-2025

Tests conducted by Embryotools

Quality Control Department
ASTEC CO., LTD.

