



ASTEC CO., LTD.
Fukuoka, Japan, 811-2207
<http://www.astec-bio.com/global/>
Email: astec-japan@astec-bio.com
Tel: +81929355666

Certificate of Analysis

Product ID : suregrip 35mm Dish
Product Code : AID-35
Lot # : 250301
Expiry date : 2028-3-7
Storage Conditions : 3-30°C, Dry, No direct sunlight

Analysis	Ref#	Results	Date of Completion
Mouse Embryo Assay80% of one cell stage control embryos to develop to expanded blastocyst within 96 hours	MEA.024.1596.2025	Pass	05-18-2025
Proven non-toxic by LAL test. Pass criteria <0.03 EU/device	LAL.024.1597.2025	Pass	05-09-2025
Sperm Motility Assay: >0.75 sperm motility by 8 hours post incubation and >0.8 sperm survival by 24 hours post incubation	SSMA.024.1598.2025	Pass	05-29-2025

Tests conducted by Embryotools

Quality Control Department
ASTEC CO., LTD.

