

Certificate of Analysis

Product ID : suregrip ICSI Dish
Product Code : AID-ICSI
Lot # : 250301
Expiry date : 2028-3-7
Storage Conditions : 3-30°C, Dry, No direct sunlight

| Analysis | Ref# | Results | Date of Completion |
|---|--------------------|---------|--------------------|
| Mouse Embryo Assay80% of one cell stage control embryos to develop to expanded blastocyst within 96 hours | MEA.024.1593.2025 | Pass | 05-18-2025 |
| Proven non-toxic by LAL test. Pass criteria <0.03 EU/device | LAL.024.1594.2025 | Pass | 05-09-2025 |
| Sperm Motility Assay: >0.75 sperm motility by 8 hours post incubation and >0.8 sperm survival by 24 hours post incubation | SSMA.024.1595.2025 | Pass | 05-29-2025 |

Tests conducted by Embryotools

Quality Control Department
ASTEC CO., LTD.

