



ASTEC CO., LTD.  
Fukuoka, Japan, 811-2207  
<http://www.astec-bio.com/global/>  
Email: [astec-japan@astec-bio.com](mailto:astec-japan@astec-bio.com)  
Tel: +81929355666

## Certificate of Analysis

Product ID : suregrip 4Well Dish  
Product Code : AID-4W  
Lot # : 250401  
Expiry date : 2028-3-15  
Storage Conditions : 3-30°C, Dry, No direct sunlight

| Analysis                                                                                                                  | Ref#               | Results | Date of Completion |
|---------------------------------------------------------------------------------------------------------------------------|--------------------|---------|--------------------|
| Mouse Embryo Assay80% of one cell stage control embryos to develop to expanded blastocyst within 96 hours                 | MEA.024.1572.2025  | Pass    | 05-17-2025         |
| Proven non-toxic by LAL test. Pass criteria <0.03 EU/device                                                               | LAL.024.1573.2025  | Pass    | 05-09-2025         |
| Sperm Motility Assay: >0.75 sperm motility by 8 hours post incubation and >0.8 sperm survival by 24 hours post incubation | SSMA.024.1574.2025 | Pass    | 05-21-2025         |

Tests conducted by Embryotools

Quality Control Department  
ASTEC CO., LTD.

