



ASTECCO., LTD.  
Fukuoka, Japan, 811-2207  
<http://www.astec-bio.com/global/>  
Email: [astec-japan@astec-bio.com](mailto:astec-japan@astec-bio.com)  
Tel: +81929355666

## Certificate of Analysis

Product ID : Pasteur Pipette 1mL  
Product Code : ATP-01  
Lot # : 202311  
Expiry date : 2026-11  
Storage Conditions : 3-30°C, Dry, No direct sunlight

Analysis	Ref#	Results	Date of Completion
Mouse Embryo Assay 80% of one cell stage control embryos to develop to expanded blastocyst within 96 hours.	MEA.022.333.2024-R1	Pass	03-02-2024
Proven non-toxic by LAL test. Pass criteria <0.03 EU/device	LAL.022.352.2024	Pass	02-14-2024
Proved stable human sperm motility: $\geq 75\%$ sperm motility after 24 hours	S7067-0224ACL	Pass	02-14-2024

Tests conducted by Embryotech / Embryotools

Quality Control Department  
ASTECCO., LTD.

