



ASTECCO., LTD.
Fukuoka, Japan, 811-2207
<http://www.astec-bio.com/global/>
Email: astec-japan@astec-bio.com
Tel: +81929355666

Certificate of Analysis

Product ID : suregrip ICSI Dish
Product Code : AID-ICSI
Lot # : 231201
Expiry date : 2026-12
Storage Conditions : 3-30°C, Dry, No direct sunlight

Analysis	Ref#	Results	Date of Completion
Mouse Embryo Assay 80% of one cell stage control embryos to develop to expanded blastocyst within 96 hours.	MEA.022.337.2024	Pass	02-24-2024
Proven non-toxic by LAL test. Pass criteria <0.03 EU/device	LAL.022.343.2024	Pass	02-16-2024
Proved stable human sperm motility: $\geq 75\%$ sperm motility after 24 hours	S7067-0224ACL	Pass	02-14-2024

Tests conducted by Embryotech / Embryotools

Quality Control Department
ASTECCO., LTD.

