

Certificate of Analysis

Product ID : suregrip ICSI Dish
 Product Code : AID-ICSI
 Lot # : 240701
 Expiry date : 2027-6-30
 Storage Conditions : 3-30°C, Dry, No direct sunlight

Analysis	Ref#	Results	Date of Completion
Mouse Embryo Assay 80% of one cell stage control embryos to develop to expanded blastocyst within 96 hours.	MEA.023.040.2024	Pass	08-18-2024
Proven non-toxic by LAL test. Pass criteria <0.03 EU/device	LAL.023.050.2024	Pass	08-16-2024
Sperm Motility Assay: >0.75 sperm motility by 8 hours post incubation and >0.8 sperm survival by 24 hours post incubation	SMA.023.089.2024	Pass	08-21-2024

Tests conducted by Embryotools

Quality Control Department
ASTEC CO., LTD.

